



Contribution ID: 71

Type: poster

## Angular control of Advanced Virgo suspended benches

*Tuesday, 24 May 2016 18:00 (0 minutes)*

Most of the Advanced Virgo light sensing devices are located on 5 in vacuum suspended benches. The benches are suspended from a triple stage mSAS suspension, and the angular degrees of freedom are sensed and controlled from the ground using LVDTs and coil-magnet actuators. The angular control has stringent requirements 30 nrad rms and  $3\text{e-}15$  rad/rtHz above 10 Hz. After introducing the system design, we present the first in air results of controlling the fully integrated system and striving to achieve these requirements.

**Primary author:** Dr WAS, Michal (LAPP/CNRS)

**Presenter:** Dr WAS, Michal (LAPP/CNRS)

**Session Classification:** Poster Session