







Temp test

Luigi Longo Università del Salento Dipartimento di Matematica e Fisica "Ennio De Giorgi" and Istituto Nazionale di Fisica Nucleare - Sezione di Lecce

Temperature Test: Experimental Setup

Chambers considered:

T1, T2 and T5

HV:

520V for the resistive strips 300V for the drift panel


Pt100

Radiactive sources:

Fe55

Flux:


1I/h for T1,T2,T5




Gas Distribution Setup

Temperature Test: Data taking procedure

- Three heating and cooling cycles for T1 and T5:
 - 1st cycle:
 - Sparks rate measured for both the chambers while the temperature (T) increases (RISE PHASE1) or decreases (DROP PHASE1) [T5 on top]
 - 2nd cycle:
 - > Sparks rate measured for both the chambers while the temperature increases (**RISE PHASE2**) or decreases (**DROP PHASE2**)
 - Iron 55 spectrum measured using T5 chambers as function of the temperature [T5 on top]
 - 3rd cycle:
 - Sparks rate measured for both the chambers while the temperature increases (RISE PHASE3), is stable (PLATEAU PHASE3) or decreases (DROP PHASE3)
 - Iron 55 spectrum measured using T1 chambers during RISE PAHSE3 and DROP PAHSE3 [T1 on top]




Temperature Test: 1st cycle - RISE PHASE1




NB: T_{Box} & T_{gas} have more or less the same behaviour; less of 1 degree of different (see backup)


Temperature Test: 1st cycle - RISE PHASE1



Temperature Test: 1st cycle - DROP PHASE1





!!! T1: too few sparks to see a trend



T1 behaviour similar to what was observed in the 1st cycle


Temperature Test: 2st cycle - Spikes RISE PHASE2 plots







¿¿ Could be that an hysteresy effect is present??

Temperature Test: 2st cycle - RISE PHASE2 Gain Vs Temp



Temperature Test:2nd cycle - DROP PHASE2







It seems that there is no hysteresis effect if one concentrates only on 2nd cycle

BUT

BEHAVIOUR DIFFERENT FROM 1ST CYCLE

Temperature Test: 2st cycle - DROP PHASE2 Gain Vs Temp






!!! For the DROP and RISE phase the gain increases in a quite linear way BUT
THE 2 CURVES ARE NOT OVERLAPPED → HYSTERESIS EFFECT??

Temperature Test: 3st cycle


Remarks:

- T1 on top of T5
- Iron55 spectrum measured from T1




!!! the effect was not been observed yet (see backup slides)


Temperature Test: 3st cycle




!!! One should assume that in the previous cycles T1 was at lower temperature wrt T5




Temperature Test: 3st cycle - RISE PHASE3




Temperature Test: 3st cycle - RISE PHASE3 Gain Vs Temp



Temperature Test: 3st cycle - PLATEAU PHASE3




Temperature Test: 3st cycle - PLATEAU PHASE3



Same T in these regions but the spark amplitude is higher in the second period wrt the previous one

Temperature Test: 3st cycle - PLATEAU PHASE3




Remarks:

Resolution of 1 s


It seems to there be a CORRELATION
BETWEEN T1 AND T5 RATE
BUT
it could be related to the fact that
there is a BAD RESOLUTION


If the same time sparks are removed an anti-correlation appears



Temperature Test:3nd cycle - DROP PHASE3


T5 continues to have an high-rate even if the T decreases






T1 rate decreases to zero decreasing the T

Temperature Test: 3nd cycle – DROP PHASE3 Gain Vs Temp





!!! For the DROP and RISE phase the gain increases in a quite linear way BUT
THE 2 CURVES ARE NOT OVERLAPPED AT HIGH & LOW T

Temperature Test: Conclusion


- Sparks are related to the high temperature, but to the gas temperature, to the chamber temperature or to both?
- In general the gain increases with the temperature but some hysteresis effects are present
- The correlation between the T1 and T5 rate is real?

- Next step:
 - Repeat the measurements increasing only gas or chamber temperature
 - Since the measurement takes a lot of time, next time the atmospheric pressure will be taken into account.


Temperature Test

Backup


Temperature Test: Tgas-TBox 1st Cycle




Temperature Test: Tgas-TBox 1st Cycle



Temperature Test: Tgas-TBox 2st Cycle



Temperature Test: Tgas-TBox 2st Cycle



Temperature Test: Tgas-TBox 2st Cycle

