

Implementation and Performance of FORE-FBP for Brain Imaging on the GE SIGNA PET/MR

Tuesday, 22 May 2018 16:15 (1h 30m)

Fourier rebinning combined with 2D filtered backprojection (FORE-FBP) was implemented for the GE SIGNA PET/MR. The algorithm was designed and tested for brain imaging. Spatial resolution, image uniformity and signal-to-noise ratio (SNR) were assessed for a range of maximum ring differences (ΔZ). Results were compared to those for ordered subsets expectation maximum (OSEM) and time-of-flight OSEM (TOF-OSEM). Nearest neighbor axial interpolation provided the highest FORE-FBP NEMA spatial resolution ($4.4 \times 4.0 \times 4.0$ mm at 1 cm radius for maximum $\Delta Z=44$), which was inferior to that for OSEM/TOF-OSEM ($3.4 \times 2.1 \times 1.5$ mm). Axial resolution for FORE-FBP degraded by < 0.4 mm for $\Delta Z=44$ compared to $\Delta Z=27$ for radii < 10 cm. Two cylindrical phantoms of disparate internal diameters (2.2 and 19.7 cm) indicated that FORE-FBP image uniformity in the axial direction in the central axial 15 cm reached optimal levels for $\Delta Z=27$. Image SNR for FORE-FBP improved with increasing ΔZ , with linear interpolation providing the lowest image variability. Image uniformity and SNR within the central axial 15 cm for the NEMA NU 2-1994 phantom were superior for OSEM/TOF-OSEM compared to FORE-FBP by 30% and 10-25%, respectively.

Primary author: Dr MANAVAKI, Roido (University of Cambridge)

Co-authors: Dr JANSEN, Floris (GE Healthcare); FRYER, Tim (University of Cambridge); Dr DELLER, Timothy (GE Healthcare); Dr HONG, Young (University of Cambridge)

Presenter: FRYER, Tim (University of Cambridge)

Session Classification: Session 8 - Poster Session I