

First Measurements with the PETsys TOFPET2 ASIC Evaluation Kit

Wednesday, 23 May 2018 12:00 (20 minutes)

We present results obtained with the PETsys TOFPET2 ASIC Evaluation Kit.

The TOFPET2 ASIC is a time-of-flight capable 64-channel ASIC and is a promising candidate to build high-performance TOF-PET systems.

For a first characterization, we used two KETEK-PM3325-WB SiPMs each equipped with a 3x3x5 mm³ LYSO scintillation crystal.

For the measurements presented in this work, we changed the discriminator threshold $t_1 = 5 - 30$ in steps of 5 and the bias voltage $V_{\text{bias}} = 29 \text{ V} - 35 \text{ V}$ in steps of 0.25 V and acquired data during 60 s.

For all measurements, we performed an energy calibration including a correction for saturation.

We evaluated the energy resolution, the coincidence resolving time (CRT) and the coincidence rate.

At the optimal operating point, we obtained an energy resolution of about 9-7% FWHM, a CRT of approximately 210 ps FWHM and 395 ps FWTM, the coincidence rate showed only small variations of about 5% in the evaluated parameter range.

Primary author: SCHUG, David (Department of Physics of Molecular Imaging Systems, Institute for Experimental Molecular Imaging, RWTH Aachen University, Aachen, Germany)

Co-authors: Dr WEISSLER, Bjoern (Department of Physics of Molecular Imaging Systems, Institute for Experimental Molecular Imaging, RWTH Aachen University, Aachen, Germany); Dr GEBHARDT, Pierre (Department of Physics of Molecular Imaging Systems, Institute for Experimental Molecular Imaging, RWTH Aachen University, Aachen, Germany); Prof. SCHULZ, Volkmar (Department of Physics of Molecular Imaging Systems, Institute for Experimental Molecular Imaging, RWTH Aachen University, Aachen, Germany)

Presenter: SCHUG, David (Department of Physics of Molecular Imaging Systems, Institute for Experimental Molecular Imaging, RWTH Aachen University, Aachen, Germany)

Session Classification: Session 11 - Instrumentation: detectors and electronics