

Channeling 2018



Contribution ID: 34

Type: Poster

Ultrashort Optical Parametric Amplifier and Oscillator up to the Near-Infrared

Wednesday, 26 September 2018 18:40 (1 hour)

Coherent light sources in spectral regions which are inaccessible to lasers has been an important issue for more than five decades. Despite tremendous progress in laser technology over the last five decades, substantial portions of the optical spectrum from the ultraviolet (UV) to the infrared (IR) still remain inaccessible to conventional laser sources, at time scales ranging from continuous to almost single-cycle pulses.

This limitation arises from the limited gain bandwidth of the active medium, which defined the operating spectral region of the laser. This directly limits the application of such devices, while also placing a boundary on ultrashort pulse generation, which requires very broad bandwidths.

In this context, coherent optical sources based on nonlinear conversion, with femtosecond pulse duration and wide tunability, are rapidly emerging. They are extremely versatile and of considerable interest for a wide range of scientific and technological applications in different areas.

Our goal is to investigate the generation of ultrashort laser pulses in the near infrared region between 900 and 1000 nm, which is of interest for current large-scale laser projects based on optical parametric chirped pulse amplification. For this, we will design and develop an innovative sub-20 fs tunable optical parametric oscillators (OPO) optimized at a chosen central wavelength and a μJ level double stage optical parametric amplification (OPA) within the new beamline project in Vulcan for betatron imaging.

The OPO systems will be implemented at two existing laser systems: the Laboratory for Intense Lasers (L2I, IST), with the goal of generating a seed pulse for a new ultrashort amplifier; and the Vulcan laser system (CLF, RAL), with the goal of generating an off-harmonic femtosecond probe line to upgrade the system. Their requirements are similar ensuring that the same technology can be applied, with suitable modification to the spectral parameters.

Primary author: Mr GALLETTI, Mario (Istituto Superior Tecnico, Universidade de Lisboa)

Co-authors: Prof. FIGUEIRA, Goncalo (Istituto Superior Tecnico); Dr PIRES, Hugo (Istituto Superior Tecnico); Dr GALIMBERTI, Marco (Science and Technology Facilities Council); Dr OLIVEIRA, Pedro (Central Laser Facility, RAL)

Presenter: Mr GALLETTI, Mario (Istituto Superior Tecnico, Universidade de Lisboa)

Session Classification: PS2 - Poster session