



Contribution ID: 42

Type: **Invited Presentation**

High-spatial resolution neutron imaging by using current-biased kinetic inductance detector

Thursday, July 25, 2019 11:45 AM (15 minutes)

We developed a neutron transmission imager based on a superconducting current-biased kinetic inductance detector (CB-KID). The CB-KID comprises X and Y meanderlines and a ^{10}B conversion layer for neutrons. A ^4He or ^7Li ion from the $^{10}\text{B}(n, \alpha)^7\text{Li}$ reaction creates two hot spots in both the X and Y meanders. A pair of electromagnetic-wave pulses of opposite polarities propagate toward the ends of meanderlines[1]. The position of the nuclear reaction point can be evaluated from a difference in arrival timestamps of the two pulses at the two ends. We used a set of analog signal discriminators with fixed thresholds and a time-to-digital converter (TDC) with 1-ns time resolution to recover the signals from 25-Hz pulsed neutrons of J-PARC. The energy-integrated spatial resolution reached $22\ \mu\text{m}$ [2]. Further improvements in spatial resolution can be achieved by using a pair of CB-KIDs to compensate the randomness of the emitted direction of light ions from the ^{10}B reaction. PHITS (Particle and Heavy Ion Transport code System) is a Monte Carlo particle transport simulation code developed to deal with the transport of all particles over wide energy ranges, using several nuclear reaction models and nuclear data libraries[3]. PHITS simulations demonstrated that the neutron imaging can be enhanced appreciably with two CB-KIDs.

This work is supported by Grant-in-Aid for Scientific Research (A) No.16H02450 from JSPS. The devices were fabricated in the clean room for analog-digital superconductivity (CRAVITY). This work is supported of MLF program of J-PARC (Proposal No. 2016B0012, 2017A0011, 2017B0014, 2018A0109, No. 2018P0201, No. 2019A0004).

1. T. Koyama, T. Ishida, J. Phys. Conf. Ser. **1054** 012055 (2018).
2. H. Shishido, Y. Miki, H. Yamaguchi, Y. Iizawa, Vu T. Dang, K. M. Kojima, T. Koyama, K. Oikawa, M. Harada, S. Miyajima, M. Hidaka, T. Oku, K. Soyama, S. Y. Suzuki, T. Ishida, Phys. Rev. Appl. **10** 044044 (2018).
3. T. Sato et al, J. Nucl. Sci. Tech. **10**, 684-690 (2018).

Less than 5 years of experience since completion of Ph.D

N

Student (Ph.D., M.Sc. or B.Sc.)

N

Primary author: Prof. ISHIDA, Takekazu (Division of Quantum and Radiation Engineering, Osaka Prefecture University)

Co-authors: Dr VU, The Dang (Materials and Life Science Division, J-PARC Center, Japan Atomic Energy Agency,); Prof. SHISHIDO, Hiroaki (Department of Physics and Electronics, Osaka Prefecture University); Mr NISHIMURA, Kazuma (Department of Physics and Electronics, Osaka Prefecture University); Prof. KOJIMA, Kenji M (Centre for Molecular and Materials Science, TRIUMF); Dr OIKAWA, Kenich (Materials and Life Science Division, J-PARC Center, Japan Atomic Energy Agency); Dr HARADA, Masahide (Materials and Life Science Divi-

sion, J-PARC Center, Japan Atomic Energy Agency); Dr MIYAJIMA, Shigeyuki (Advanced ICT Research Institute, National Institute of Information and Communications Technology); Dr HIDAKA, Mutsuo (National Institute of Advanced Industrial Science and Technology); Dr OKU, Takayuki (Materials and Life Science Division, J-PARC Center, Japan Atomic Energy Agency); Dr SOYAMA, Kazuhiko (Materials and Life Science Division, J-PARC Center, Japan Atomic Energy Agency); Dr AIZAWA, Kazuya (Materials and Life Science Division, J-PARC Center, Japan Atomic Energy Agency); Dr KOYAMA, Tomio (Division of Quantum and Radiation Engineering, Osaka Prefecture University.); Dr MALINS, Alex (Center for Computational Science & e-Systems, Japan Atomic Energy Agency); Dr MACHIDA, Masahiko (Center for Computational Science & e-Systems, Japan Atomic Energy Agency)

Presenter: Prof. ISHIDA, Takekazu (Division of Quantum and Radiation Engineering, Osaka Prefecture University)

Session Classification: Orals LM 004

Track Classification: Low Temperature Detector Applications