



Contribution ID: 218

Type: **Poster**

## Development of low threshold detectors for light dark matter detection

*Tuesday, July 23, 2019 6:45 PM (15 minutes)*

We developed a 5x5x5 mm<sup>3</sup> crystal detector with an MMC readout. The detector was designed to achieve low energy threshold for direct detection of low mass dark matter. A pure CaF<sub>2</sub> crystal was adopted as a target. This absorber crystal had a strong thermal contact to a metallic magnetic calorimeter (MMC) sensor via thin gold film evaporated on its surface. The MMC sensor and the gold film were directly bonded together using cold diffusion welding. We will present the result of the detector performance together with the detector model of the heat flow.

### Less than 5 years of experience since completion of Ph.D

Y

### Student (Ph.D., M.Sc. or B.Sc.)

Y

**Primary author:** KIM, Hyelim

**Co-authors:** Dr KIM, SG (Center for Underground Physics, Institute for Basic Science); Ms KIM, SR (Center for Underground Physics, Institute for Basic Science); Ms LEE, HJ (Center for Underground Physics, Institute for Basic Science); JEON, JinA; LEE, Minkyu; KIM, Young-Hamb (Institute for Basic Science)

**Presenter:** KIM, Hyelim

**Session Classification:** Poster session

**Track Classification:** Low Temperature Detector Development and Physics