



Contribution ID: 276

Type: Poster

A New Measurement of the 60 keV Transition in Am-241 Decays using Metallic Magnetic Calorimeters

Tuesday, July 23, 2019 6:45 PM (15 minutes)

The 60 keV transition in Am-241 decay is one of the most important calibration standards for low energy gamma-rays. The current literature value of 59.5409(1) keV is based on measurements with high-purity Ge detectors and a Tb-161 reference source in 1993, and its 0.1 eV uncertainty gives it significant weight for cryogenic detector calibration. We have re-measured the energy of this transition in Am-241 decays with metallic magnetic calorimeter (MMC) gamma detectors with an energy resolution of 80 eV and demonstrated high linearity and reproducibility. For calibration, we have made a Yb-169 source, whose gamma emissions are known extremely accurately from measurements using crystal spectrometers, through Tm-169(d,2n)Yb-169 at the 88" Cyclotron. We will discuss statistical and systematic uncertainties of the measurements and provide a preliminary recommendation for an improved value of the Am-241 gamma-ray energy.

Less than 5 years of experience since completion of Ph.D

Y

Student (Ph.D., M.Sc. or B.Sc.)

N

Primary authors: KIM, Geon-Bo (Lawrence Livermore National Laboratory); Prof. BOYD, Stephen (University of New Mexico); Dr CANTOR, Robin (STAR Cryoelectronics); Prof. BERNSTEIN, Lee (Lawrence Berkeley National Laboratory); Dr FRIEDRICH, Stephan (Lawrence Livermore National Laboratory)

Presenter: KIM, Geon-Bo (Lawrence Livermore National Laboratory)

Session Classification: Poster session

Track Classification: Low Temperature Detector Applications