



Contribution ID: 33

Type: **Poster**

## **An array scalable zero-bias far-IR detector with noise thermometry readout**

*Tuesday, July 23, 2019 6:45 PM (15 minutes)*

We report on a new development effort to achieve an array of ultra-sensitive ( $NEP < 1E-20$  W/sqrt(Hz)) far-IR detectors for applications in spectrometers on Origins Space Telescope (OST) or similar low-background platforms. The detector uses a submicron-size hot-electron bolometer (HEB) sensor made from normal metal (non-superconducting Ti) coupled to a planar microantenna. The detector does not require any bias (dc or rf). The Johnson Noise Thermometry using a quantum noise limited microwave amplifier (LNA) allows for the direct read of an increase of the electron temperature caused by the absorbed far-IR radiation. At 50 mK, the NEP is less than  $1E-20$  W/sqrt(Hz) is expected whereas the dynamic range is 60-100 dB. Multiplexing of a 1000-pixel array is feasible using a single LNA with a bank of narrowband bandpass filters for channel multiplexing.

In this paper, we will present an initial experimental study of the electrical NEP in a  $1\mu\text{m} \times 1\mu\text{m}$  detector. A set of superconducting narrow band-pass ( $Q = 100-1000$ ) and low-pass filters defines the readout bandwidth around the center frequency of 1.5 GHz. A commercial HEMT LNA with the noise temperature  $T_A \approx 1$  K largely determines the system sensitivity ( $NEP \approx 1E-19$  W/sqrt(Hz) @ 50 mK). Electrical NEP is measured by sending a dc current through the device and measuring a change of the output noise power caused by the heating. Switching to a quantum noise limited parametric kinetic inductance amplifier will allow us to reach an NEP close to  $1E-20$  W/sqrt(Hz). The next phase of this work will be using much smaller HEB devices (e.g.,  $0.5\mu\text{m} \times 0.25\mu\text{m}$ ) where  $NEP = 3E-21$  W/sqrt(Hz) is predicted. Because of the very high dynamic range and optical saturating power, various additional higher background or/and higher operating temperature applications of such a sensor are envisioned.

### **Less than 5 years of experience since completion of Ph.D**

N

### **Student (Ph.D., M.Sc. or B.Sc.)**

N

**Author:** Dr KARASIK, Boris (Jet Propulsion Laboratory)

**Co-authors:** Dr DAY, Peter (Jet Propulsion Laboratory); Dr SKALARE, Anders (Jet Propulsion Laboratory)

**Presenter:** Dr KARASIK, Boris (Jet Propulsion Laboratory)

**Session Classification:** Poster session

**Track Classification:** Low Temperature Detector Development and Physics