



Contribution ID: 88

Type: Poster

Hyperspectral X-ray Imaging

Thursday, 25 July 2019 18:45 (15 minutes)

We are developing a chemical imaging capability (“Hyperspectral X-ray Imaging”) for microscopic samples based on ultra-high-resolution x-ray emission spectroscopy with large transition-edge sensor microcalorimeter arrays in the scanning electron microscope. By combining microcalorimeter arrays with hundreds of pixels, high-bandwidth microwave frequency-division multiplexing, and fast digital electronics for near real-time data processing, our goal is to enable practical chemical speciation analysis using small-laboratory instrumentation rather than synchrotron beamlines. Our focus is on mapping the detailed chemical form of microscopic particles containing materials from the nuclear fuel cycle. Their detailed chemical form is a crucial link to material origin, history, and behavior in the environment. In combination with developing the instrumentation to obtain high-quality x-ray emission spectra on such samples, we are working to develop a validated theory capability to interpret fine structure in the spectra and better understand fundamental properties of actinide chemical bonding. We will present our approach to developing the Hyperspectral X-ray Imaging capability, recent results from a 128-pixel microcalorimeter array at LANL and the 16-pixel STAR Cryoelectronics MICA-1600 spectrometer, and the path to high-throughput chemical mapping.

Less than 5 years of experience since completion of Ph.D

N

Student (Ph.D., M.Sc. or B.Sc.)

N

Primary author: CROCE, Mark (Los Alamos National Laboratory, USA)

Co-authors: BAKER, Zachary (Los Alamos National Laboratory); BATISTA, Enrique (Los Alamos National Laboratory); CAFFREY, Michael (Los Alamos National Laboratory); FONTES, Christopher (Los Alamos National Laboratory); KOEHLER, Katrina (Los Alamos National Laboratory); KOSSMANN, Shannon (Los Alamos National Laboratory); KOZIMOR, Stosh (Los Alamos National Laboratory); MCINTOSH, Kathryn (Los Alamos National Laboratory); RABIN, Michael (Los Alamos National Laboratory); RENCK, Bryce (Los Alamos National Laboratory); WAGNER, Gregory (Los Alamos National Laboratory); YANG, Ping (Los Alamos National Laboratory); YOHO, Michael (Los Alamos National Laboratory); WILKERSON, Marianne (Los Alamos National Laboratory); BECKER, Daniel (National Institute of Standards and Technology); BENNETT, Douglas (National Institute of Standards and Technology); GARD, Johnathon (National Institute of Standards and Technology); IMREK, Jozsef (National Institute of Standards and Technology); MATES, John A.B.; MORGAN, Kelsey (National Institute of Standards and Technology); O’NEIL, Galen (National Institute of Standards and Technology); REINTSEMA, Carl (National Institute of Standards and Technology); SCHMIDT, Dan (National Institute of Standards and Technology); SWETZ, Daniel (National Institute of Standards and Technology); WESSELS, Abigail (National Institute of Standards and Technology); ULLOM, Joel (National Institute of Standards and Technology); CANTOR, Robin (STAR Cryoelectronics); HALL, Ad (STAR Cryoelectronics); CARVER, Travis (STAR Cryoelectronics)

Presenter: CROCE, Mark (Los Alamos National Laboratory, USA)

Session Classification: Poster session

Track Classification: Low Temperature Detector Applications