


# Muon g-2 laser overview

*Andrea Fioretti, CNR-INO and INFN, Pisa, Italy*


*on behalf of the g-2 italian collaboration*

*MUSE General Meeting, Frascati 23-25 October 2019*



- Introduction
- Laser *distribution* system
- Laser *monitoring* system
- Control Electronics and data readout
- *Operation modes and gain calibration*
- Conclusions

# The laser calibration system



## Idea:


- Send trains of laser pulses on known intensity synchronously on all calorimeters' channels (24X54=1296 SiPMs)

## Goals:

- Calibration of the of the SiPMs gain: short term (in fill – 700  $\mu$ s, gain sag) at  $10^{-4}$  level; long term (bias and temperature variations) gain function at  $10^{-3}$  level
- Synchronization signals of the 1296 traces: sync pulse before muon
- Debugging of Calorimeters and Data Acquisition System (flight simulator mode)
- Gain equalization of the SiPMs response (photoelectrons/photons response)

# Outline



---



- Introduction
- Laser *distribution* system
- Laser *monitoring* system
- Control Electronics and data readout
- *Operation modes and gain calibration*
- Conclusions





# The laser system





6 Picoquant Laser diodes @405nm, 600ps, 1nJ/pulse, 0-40 MHz rep. rate

# The laser distribution system







# The laser system









# The diffusing system for each calorimeter




Crystal




Engineered diffuser

Launching fiber

Delrin panel





Monitor fiber





Panel holding 54 reflecting prisms, one in front of each crystal

# A special system: the Double-Pulse setup



**Double Pulse  
OFF**

# A special system: the Double-Pulse setup





## Double Pulse ON

It is used to study  
the Gain Sag of  
SIPMs  
(**pileup** and **splash**)


Delay generator:  
(SRS DG645)  
from **ps** to **ms**




# Double-pulse mode ON





# Double-pulse mode ON




Double Pulse,  $\Delta t = 4$  ns



Double Pulse,  $\Delta t = 32$  ns




**Short-Term  
Double Pulse  
(see **Paolo  
GIROTTI's** talk)**








# Double-pulse mode ON




Double Pulse,  $\Delta t = 4$  ns



Double Pulse,  $\Delta t = 32$  ns




**Short-Term  
Double Pulse  
(see Paolo  
GIROTTI's talk)**




Fixed  
mirrors

**Long-Term Double  
Pulse  
(see Elia  
BOTTALICO's talk)**





# Synchronization of *T0* and *Fiber Harps*



Installation of the **optics** and **2 new optical fibers** to send laser signals to **Fiber Harps**






# Synchronization of *T0* and *Fiber Harps*



Installation of the **optics** and **2 new optical fibers** to send laser signals to **Fiber Harps**

# Synchronization of *T0* and *Fiber Harps*



Engineered diffuser




Fiber bundle



Installation of the **optics** and **2 new optical fibers** to send laser signals to **Fiber Harps**



# Synchronization of *T0* and *Fiber Harps*



Engineered diffuser


Fiber bundle



Installation of the **optics** and **2 new optical fibers** to send laser signals to **Fiber Harps**

**T0 detectors** have been connected to **spare fibers from the diffusers** of 2 calorimeters

# Laser energy in the distribution system




Position in distribution/monitoring chain	Energy per laser pulse	notes
Laser head	700 pJ	Only from one laser head From now on values are with filter n. 6 (standard working mode)
After Source Monitor	490 pJ	
Onto Source Monitor	210 pJ	
Towards Fiber Harps detectors	70 pJ	
After filter wheel	170 pJ	
Before each launching fiber	40 pJ	Diffuser-bundle at 40 mm working distance
Onto each diffuser	30 pJ	
Onto each front panel's fiber	30 fJ	
Onto each PbF <sub>2</sub> crystal	18 fJ	
Onto each SiPM	4 fJ	
Onto LM from diffuser	4 – 6 fJ	
Onto T0 detector	4 – 6 fJ	
Onto Fiber Harps	5 – 10 fJ	

**Table 1:** Laser energy per single pulse at various points of the laser distribution and monitoring systems. These values are only indicative and may vary by 5-10% at each step going from one distribution line to another.





- Introduction
- Laser *distribution* system
- Laser ***monitoring*** system
- Control Electronics and data readout
- *Operation modes and gain calibration*
- Conclusions

# The source monitor




- It takes 30% of the laser beam
- It is based on an **integrating sphere** illuminating **2 PIN photodiodes**, **1 PMT+ reference pulser** (Am/Nal source), **one fiber minibundle** for LM
- It provides an early reference signals for the PMTs of the local monitor.







# The source monitor





- It takes 30% of the laser beam
- It is based on an **integrating sphere** illuminating **2 PIN photodiodes**, **1 PMT+ reference pulser** (Am/Nal source), **one fiber minibundle** for LM
- It provides an early reference signals for the PMTs of the local monitor.



# The source monitor







- It takes 30% of the laser beam
- It is based on an **integrating sphere** illuminating **2 PIN photodiodes**, **1 PMT+ reference pulser** (Am/Nal source), **one fiber minibundle** for LM
- It provides an early reference signals for the PMTs of the local monitor.



**Laser and Am/Nal source on PMT**  
DAQ in Asynchronous Mode





# Performance of the Source Monitor



Statistical distributions of laser signals from (a) the PMT and (b) one of the PIN diodes, as measured at the input to the DAQ

(c) The variation of the ratio between the two PIN diode signals for SM1. Each point is an average of 3000 ratios and the two dashed lines indicate the  $10^{-4}$  stability limit over the 80-minutes acquisition period.

# Stability of SM versus Temperature








Stability of SM versus temperature changes in a two-week period.

A linear correction is applied to the data to compensate for the temperature effect.


**The final stability is better than 0.2% for the single PIN**

# Stability of SM versus Temperature






Stability of SM versus temperature changes in a two-week period.  
A linear correction is applied to the data to compensate for the temperature effect.  
**The final stability is better than 0.2% for the single PIN and  $10^{-4}$  for the ratio.**



# The local monitor




- 24 + 24 PMTs (TWO fore ach calorimeter)
- Each PMT receives **an early pulse** from the SM integrating sphere
- Each PMT receives **a pulse from a calorimeter** (through an optical fiber)
- **The ratio of the two pulse amplitudes monitors the optical transmission from the SM to each calorimeter**



# Performance of the Local Monitor





Ratio of amplitudes of the delayed LM peak to the respective SM peak for one particular LM channel. It is below the per mille level.



- Introduction
- Laser *distribution* system
- Laser *monitoring* system
- Control Electronics and data readout
- *Operation modes and gain calibration*
- Conclusions





# The control electronics and data readout





Picture of the **Laser Control Board**, which is implemented by a hybrid platform hosting a Spartan6 FPGA board and an embedded CPU

# The control electronics and data readout




# The control electronics and data readout



Picture of the **Laser Control Board**, which is implemented by a hybrid platform hosting a Spartan6 FPGA board and an embedded CPU

Schematics of the locally implemented data acquisition system, based on multiple crates.










- Introduction
- Laser *distribution* system
- Laser *monitoring* system
- Control Electronics and data readout
- *Operation modes and gain calibration*
- Conclusions

# Flight simulator mode





Laser pulses are generated according to an exponential function

This function was heavily used during the commissioning phase for Calorimeters and DAQ debugging





# Flight simulator mode



A gain sag was evident already in Flight Simulator Mode  
**FSM is NOT used to compute gain corrections**




# SiPMs gain equalization




Plot of the **variance  $\sigma^2$**  versus **mean pulse-integral,  $M$** , of a distribution of fitted laser pulses on a SiPM. The different discrete mean values are obtained using a multi-step filter wheel to attenuate the light intensity

A typical distribution of pulse energies for a given laser intensity



# Double-Pulse operating mode



**It is used to study the response of the Calorimeters to a large number of particle**

Short-Term, Double-Pulse (0-100 ns, 1 ns step) analysis



**SIPM response**



see Paolo's talk

Long-Term, Double-Pulse (0-200  $\mu$ s, 1  $\mu$ s step) analysis





**Power supply response**



see Elia's talk

**Double-Pulse studies should be performed regularly (each 3 days)  
to compute gain corrections**



# Standard operation for data taking



In-Fill pulses are **prescaled** and **time-shifted** in time from fill to fill




# (old) Results for the gain correction







$10^{-4} / h$  stability demonstrated  
with mono-energetic test beam at SLAC

# (new) Results for the gain correction




- Out-of-Fill gain corrections
- In-Fill gain corrections
- Short- and Long-Term Double-Pulse studies

ALL THESE TOGETHER ARE USED TO COMPUTE THE GAIN CORRECTION TABLES



# DQM laser monitor page




Local monitor

Source monitor


relative timing  
of the 6 laser



# Conclusions



- The main purpose of the laser calibration system is to provide **calibration and time alignment** to the 1296 SiPMs and to guarantee **gain stability** to achieve the challenging precision of 20 ppb on the  $\omega_a$  measurement
- The adopted solution is based on a **triggerable diode laser system with multiple laser heads with fluctuations below the percent level**, an **optical distribution system ensuring adequate intensity and homogeneity**, and a **system for monitoring the laser system itself with a stability at the 0.01% level**.
- The laser system has been running with no main problems for the last three years, providing debugging DAQ and Calorimeters. Now it provides gain correction tables




The Muon g-2 experiment will look for deviations from the standard model by measuring how muons wobble in a magnetic field.

PARTICLE PHYSICS

## Muons' big moment

Nature, April 11<sup>th</sup> 2017

The end - *Thank you for your attention*





# Backup slides

---



# Backup





# Filter wheel calibration




High rate laser pulses (10 kHz) to each SiPM (crystal), fit with gaussian distribution.






# Filter wheel calibration






Plot of variance  $\sigma^2$  versus pulse-integral mean  $\mu$  of the distribution using a multi-step filter wheel to attenuate the light.





# Segmented PbF<sub>2</sub> Calorimeter with SiPM Readout

SiPM boards optimized to produce PMT-like pulses to exploit short pulse duration of Cherenkov crystals (relevant: pileup)




28-channel prototype calorimeter tested at SLAC



Nature, April 11<sup>th</sup> 2017

FERMILAB



The Muon g-2 experiment will look for deviations from the standard model by measuring how muons wobble in a magnetic field.

PARTICLE PHYSICS

<http://www.nature.com/news/muons-big-moment-could-fuel-new-physics-1.21811>

# Muons' big moment



# Systematics

**Pileup:** two low energy positrons fake a high energy positron  
(happens early, not late)

$$\phi_{\text{early}} \sim \phi_1 + \phi_2$$


$$\phi_{\text{late}} \sim \phi_1$$

**Gain change:**  
example: saturation (happens early, not late)

$$\phi_{\text{early}} \sim \phi_1$$

$$\phi_{\text{late}} \sim \phi_2$$

$$\Delta\omega \sim \Delta\phi$$



Design not driven by absolute performance, but relative stability early to late  
Know how well we did on these for BNL experiment. Need to do better by a factor of 4. Detector package designed to contain the tools to enable this.

## Muon $g-2$ SM test uncertainties summary

Quantity	Uncertainty $\times 10^{-11}$	$\delta a_\mu / a_\mu$ (ppb)
$\omega_a$ statistical	53	458
$\omega_a$ systematic	24	210
$\tilde{\omega}_p$ systematic	20	170
CODATA $m_\mu / m_e$	2.6	22
CODATA $\mu_p / \mu_e$	0.35	3
Electron $g$ factor, $g_e$	0.000035	0.0003
QED	0.08	0.7
EW	1	8.6
hadLBL	26	223
hadLO	33	280
hadNLO	0.9	7.6
hadNNLO	0.1	0.86

# Uncertainty of E821 $\omega_a$ measurement




## Statistical uncertainty

- fit  $N_e(E_e > E_{thr}) = N_0(E_{thr})e^{-t/\gamma\tau} [1 + A(E_{thr}) \cos(\omega_a t + \phi(E_{thr}))]$
- $\frac{\delta\omega_a}{\omega_a} = \frac{1}{\omega_a \gamma \tau_\mu} \sqrt{\frac{2}{NA^2 P^2}}$   $N$  = number of muons,  $P$  = muon polarization,  $A$  = asymmetry
  - ▶ improves with  $B$  field since  $\omega_a \propto B$
  - ▶ improves with number of muons, asymmetry, polarization
  - ▶ improves with muon momentum ( $\gamma$ )

## Systematic uncertainty

	E821 [ppb]
detector gain variation	120
event pileup	80
lost muons	90
coherent betatron oscillations	70
electric-field and pitch corrections	50

# Data Acquisition System




- Calorimeters, trackers and the laser monitoring system are read out by custom 800 MSPS waveform digitizers.
- The DAQ produces a deadtime-free record of each 700  $\mu$ s muon fill. We get 12 fills per second, providing a total data rate of 20 GB/s.
- Data from each calorimeter is processed by an NVidia Tesla K40 GPU, which processes 33M threads per event.
- Data is sorted by T-method (chopped islands) and Q-method (current integrated) data, from which timing info can be extracted.
- The DAQ software is MIDAS based







# A Temperature problem



One laser head failed due to high room temperature  
Independent conditioning needed