

RPC Ecogas Study – Updates



Amrutha Samalan


- ECOGAS1: 45% HFO, 50% CO₂, 4% iC₄H₁₀, 1% SF₆


Analysis: scans - 03 April 2019 to 30 October 2019


- ECOGAS2: 35% HFO, 60% CO₂, 4% iC₄H₁₀, 1% SF₆

Analysis: scans- 07 November 2019 to 11 March 2020

- The dark current monitored in all scans of the ecogas study is plotted (7 chambers)
- Data points corresponding to Ecogas1 and Ecogas2 is separated with a bold dotted line
- Atlas chamber was not tested with Ecogas2










Ecogas 1

CMS-KODEL-1-4-BOT

Ecogas 2








Ecogas 1


ATLAS-2-0




Backup slides

Dark current monitoring summary - February

CMS-GT-2-0-BOT




CMS-GT-2-0-TOP




Dark current monitoring summary - February

CMS-KODEL-1-4-BOT




CMS-KODEL-1-4-TOP




Dark current monitoring summary - February



ALICE-2-0



EPDT-RPC-3




Dark current monitoring summary - December & January




Dark current monitoring summary - December & January

CMS-KODEL-1-4-BOT




CMS-KODEL-1-4-TOP



Dark current monitoring summary - December & January


ALICE-2-0



December

January



EPDT-RPC-3




December

January


Variation in gas flow and dew point due to insufficient CO₂ supply (17/02/2020)




CMS-KODEL_1-4-TOP




CMS-KODEL_1-4-BOT




CMS-GT-2-0-TOP



CMS-GT-2-0-BOT






ALICE-2-0



Chamber/Gap	Dark current (µA) (@final HV point)
CMS-GT-2-0-BOT	3.44
CMS-GT-2-0-TOP	9.29
CMS-KODEL-1-4-BOT	25.33
CMS-KODEL-1-4-TOP	6.12
ALICE-2-0	32.44

- Chamber EPDT3 was removed on wednesday


Dark current monitoring-summary (February)



Dark current monitoring-summary (February)


CMS-GT-2-0-TOP

@11500 V




EPDT-RPC-3

@11500 V



CMS-GT-2-0-BOT

@11500 V



Summary

- Total gas flow and dew point decreased due insufficient supply of CO₂
- Dark current monitored during HV scan (19/02/2020) is pretty high for CMS-KODEL_1-4-BOT (25.33 μ A) gap and ALICE chamber (32.44 μ A)
- Removed chamber EPDT3 on wednesday
- CMS-GT_2-0 gap is showing unexpected behaviour in terms of dark current
- New HV scan started today @10.04AM