# APTS OpAmp Update


Arianna, Angelo, Francesco et al.

## **APTS OpAmp**

- APTS (Analogue Pixel Test Structure) is one of test structures of the first chip submission (MLR1) in TPSCo65 nm CMOS imaging process
- Detailed characterization of the smaller feature size (ALPIDE is 180 nm CMOS process sensor)
- 2 variants of the APTS chip:
  - APTS Source Follower (SF):
    - Frontend based on source follower structure
    - Robust readout but slow
  - APTS Operational Amplifier (OA):
    - High speed OPAMP buffers out frontend output
    - Study charge collection properties and timing features of the sensor
    - $50 \,\Omega$  terminating resistance on board
    - Identical frontend as APTS SF

## Pixel matrix implementation

- Matrix sizes: 6×6 pixels matrix
- **Readout**: only central 4x4 pixels are analogue read out
- **Pixel pitch**: square pixels of **10**, 15, 20, 25 μm
- Innermost 2 ×2 pixels are connected to SMA outputs to be read via oscilloscope
- External 12 pixels are read via (slow) ADCs



## **Sensor features**

Collection electrode


Additional deep implant

P\* SUBSTRATE

P implant


- Sensor based on 65nm TPSCo CMOS imaging process
- 3 different pixel flavours: •
  - Standard process ٠
  - Modified process ٠
  - Modified with gap ٠

Sensor under test in Bari •



• • STANDARD • • NWELL COLLECTION ELECTRODE PMOS \_ PWELL NWELL PWELL DEEP PWELL DEPLETED ZONE P\* EPITAXIAL LAYER

## Pixel fronted APTS OpAmp schematic version




- All operation parameters are fixed except V<sub>reset</sub> which can be optimized at varying V<sub>substrate</sub>
- $V_{substrate} = V_{p-well}$
- Charge injection system to calibrate chip
  - $V_h$  voltage generated by DAC + injection capacitance  $C_{inj}$

## **Current Studies**

- Calibration:
  - Gain calibration
  - Pulsing calibration
  - $V_h$  scan calibration
- <sup>55</sup>*Fe* spectrum measurements

## **Experimental Setup**

- MLR1 DAQ board connected via USB to DAQ machine
- Oscilloscope connected via TCP port to DAQ machine
- Power supply connected via USB to DAQ machine





Power supply Rohde&Schwarz HAMEG HMP4040







10 05 1

Oscilloscope Rohde&SchwarzRTO1044 (4 Ghz BW - 20 Gsa/s)


- Measure the baseline of the 16 pixels at varying  $V_{reset}$  (20 mV to 900 mV interval)
- Data processing and analysis produce mean and rms of *Baseline* and  $dBaseline/dV_{reset}$




- Measure the baseline of the 16 pixels at varying  $V_{reset}$  (20 mV to 900 mV interval)
- Data processing and analysis produce mean and rms of *Baseline* and *dBaseline/dV*<sub>reset</sub>




- Measure the baseline of the 16 pixels at varying  $V_{reset}$  (20 mV to 900 mV interval)
- Data processing and analysis produce mean and rms of *Baseline* and  $dBaseline/dV_{reset}$




- Measure the baseline of the 16 pixels at varying  $V_{reset}$  (20 mV to 900 mV interval)
- Data processing and analysis produce mean and rms of *Baseline* and  $dBaseline/dV_{reset}$




- Measure the baseline of the 16 pixels at varying  $V_{reset}$  (20 mV to 900 mV interval)
- Data processing and analysis produce mean and rms of *Baseline* and  $dBaseline/dV_{reset}$




- Measure the baseline of the 16 pixels at varying  $V_{reset}$  (20 mV to 900 mV interval)
- Data processing and analysis produce mean and rms of *Baseline* and *dBaseline/dV*<sub>reset</sub>



- Measure the baseline of the 16 pixels at varying  $V_{reset}$  (20 mV to 900 mV interval)
- Data processing and analysis produce mean and rms of *Baseline* and  $dBaseline/dV_{reset}$




- Measure the baseline of the 16 pixels at varying  $V_{reset}$  (20 mV to 900 mV interval)
- Data processing and analysis produce mean and rms of *Baseline* and  $dBaseline/dV_{reset}$



## **Pulsing calibration**

- The current analysis processes only the internal pixels
- Acquire test pulses produced by the proxy board ( $V_h$  DAC) and injected via injection capacitance in the pixels frontend
- $V_{reset}$  is measured where  $\frac{dBaseline}{dV_{reset}} > 0,6$  for the injected charge corresponding at **fixed**  $V_h = 1200mV$




• Analysis produces the optimal  $V_{reset}$  for each  $V_{bb}$  applied (working point)

#### **Operation Point W22AO10Pb18**

$V_{bb}(V)$	$V_{reset} (mV)$
0.0	260
-0.6	310
-1.2	350
-1.8	370
-2.4	390
-3.0	390
-3.6	400
-4.8	400

Working point configuration: maximum signal amplitude, maximum gain, minimum sighal falltime


#### **Pulsing calibration output: Signal Amplitude**



working point configuration: maximum signal amplitude, maximum gain, minimum signal falltime

V<sub>reset</sub> (mV)

### Pulsing calibration output: dV/dt 10%-90%




working point configuration: maximum signal amplitude, maximum gain, minimum signal falltime

V<sub>reset</sub> (mV)

#### Pulsing calibration output: Fall-Time 10%-90%

Falltime (ps)



working point configuration: maximum gain, maximum signal amplitude, minimum signal falltime

V<sub>reset</sub> (mV)


## $V_h$ scan calibration

- The current analysis processes only the internal pixels
- For the working points defined in the pulsing calibration, acquire test pulses produced by the proxy board (V<sub>h</sub> DAC) at varying injected charge (V<sub>h</sub> scan default range: 200 mV to 1200mV in steps of 100mV)

#### **Operation Point W22AO10Pb18**

$V_{bb}(V)$	$V_{reset} (mV)$
0.0	260
-0.6	310
-1.2	350
-1.8	370
-2.4	390
-3.0	390
-3.6	400
-4.8	400


#### $V_h$ scan calibration output: Signal Amplitude



Signal amplitude (mV)

#### *V<sub>h</sub>* scan calibration output: Fall-Time 10%-90%

Falltime (ps)




 $V_h$  (mV)

Energy (keV)	Intensity (%)	Туре
5.89875	16.57	$X_{K\alpha 1}$
5.88765	8.45	$X_{K\alpha 2}$
6.5128	3.40	$X_{K'\beta 1}$
0.6385	0.524	$X_L$
125.949	0.00000013	γ


<sup>55</sup>*Fe* Fall-Time 10%-50%

h2d\_dV\_Ft\_1050\_n h2d\_dV\_Ft\_1050\_n Entries 16802 time[ps] 80 Mean x 0.04096 900 307.2 Mean y 70 Std Dev x 0.01903 800 Std Dev y 145.9 60 700 50 600 500 40 400 30 300 20 200 10 100 00 0 0.01 0.02 0.03 0.04 0.05 0.06 0.07 0.08 amplitude [V]


h\_deltaV\_Sel\_Ch1




h\_deltaV\_Sel\_Ch2



h\_deltaV\_Sel\_Ch3



h\_deltaV\_Sel\_Ch4



h\_deltaV\_Sel\_Ch1

