VERTEX2023 - 32nd International Workshop on Vertex Detectors



Operation and performance of the NA62 GigaTracKer

EPFL

Alina Kleimenova (EPFL, Lausanne) <u>akleimen@cern.ch</u> on behalf of GTK working group: CERN, UCL Louvain, Università/INFN Ferrara, Università/INFN Torino




www.cinqueterre5.com

Sestri Levante 2023, 16th October

Outline




- Overview of the NA62 experiment
- GigaTracKer, the beam spectrometer of NA62
- Performance
- Prospects
- Summary





The NA62 experiment



~30 institutes, ~200 participants

NA62 is a fixed-target experiment at CERN SPS

Main goal: measure $\mathcal{B}(K^+ \to \pi^+ \nu \bar{\nu})$ with 10% precision using novel kaon-in-flight technique

Current SM prediction:

 $\mathcal{B}(K^+ \to \pi^+ \nu \bar{\nu}) = (8.60 \pm 0.42) \times 10^{-11}$ [arXiv:2109.11032]


Experimental values: $\mathcal{B}(K^+ \to \pi^+ \nu \bar{\nu}) = (17.3^{+11.5}_{-10.5}) \times 10^{-11}$ E949/E787[Phys. Rev D 79, 092004 (2009)] $\mathcal{B}(K^+ \to \pi^+ \nu \bar{\nu})$ $= (10.6^{+4.0}_{3.4 \ stat} \pm 0.9_{syst}) \times 10^{-11}$ NA62[JHEP06 (2021) 093]

Broader physics programme:

- Rare/forbidden kaon decays
- Searches for **exotic particles** in kaon decays and in **beam dump** mode



Detector overview



Primary beam:


- 400 GeV protons from SPS
- $\sim 3.3 \times 10^{11}$ protons per spill
- average spilllength ~4s

Secondary beam:

- 75 GeV hadron beam, $\frac{\Delta p}{n} = \sim 1\%$ (rms)
- 6% of the beam are kaons




Detector overview




Performances:

- GTK-KTAG-RICH time resolution: $\mathcal{O}(100 \text{ ps})$
- $\mathcal{O}(10^4)$ background suppression from kinematics
- $\mathcal{O}(10^7)$ muon rejection for 15 <p(π^+)< 35 GeV $\mathcal{O}(10^8)$ π^0 rejection of for E(π^0)> 40 GeV
- •



[NA62 Detector Paper, JINST 12 (2017), P05025

The beam spectrometer of NA62





Design specifications

- Design started in 2007
- Three planes of Si hybrid pixels $\rightarrow 4^{\text{th}}$ layer (GTK0) was installed in 2021 in the same vessel with GTK1
- Installed in beam pipe vacuum: $\sim 10^{-6}$ mbar
- Detectors are supposed to be replaced after 1 year at full intensity

Beam rate	750MHz-1GHz
Peak particle flux	2.0 MHz/mm ²
Peak radiation	4.5×10 ¹⁴ 1MeV n _{eq} /cm ² /200
	days
Efficiency	99%
Momentum	0.2%
resolution	
Angular resolution	16 µrad
Pixel time	<200ps RMS
resolution	
Material budget	0.5%X ₀
Detector size	60.8x27mm ²



The GigaTracKer detector



Detailed description can be found <u>JINST14 P07010</u>


16/10/2023

The pixel matrix



TDCPix

Sensor



Sensor:

- Planar n-in-p/p-in-n
- $27 \times 60.8 \,\mathrm{mm^2}$
- 200 μ m thick (0.2% X₀)
- MPV of charge per MIP: 2.4fC, ~15000e

Bump-Bonding:

- Sn-Ag bumps
- Benzocyclobutane deposited to avoid discharges


TDCPix:

- IBM 130nm CMOS technology
- 100 μ m thick (0.1% X₀) 1800 pixels of 300 × 300 μ m²
- Time resolution: <200ps
- Peaking time: 5ns
- TDC bin size: 97ps
- SEU mitigation
- FCN <250e

See more in NIMA.2023.168331

Cooling plate

- Micro-channel technology first application in HEP
- . Fabricated by CEA Leti
- . Dimensions: 70x80 mm², design thickness 210 μ m (0.2% X₀)
- 2 groups of 75 microchannels, 70x200µm²
- Liquid coolant C₆F₁₄ (~3g/s @ 3bars)
- Front-end electronics and sensor at <5°C



Sensor

TDCPix

KOVAR connectors

70um 150um 200um



Detailed description can be found <u>JINST14 P07010</u>



Electrical Integration





PCB:

- Chips, sensor and cooling plate assembly hosted in the countersink of the carrier board
- 14 layers T-shaped PCB
- 40 differential 3.2 Gb/s signals over 30cm

Wire bonding:

- TDCPix wired bonded to PCB
- Dense bonding scheme with 73 µm pitch on TDCPix (1450 bonding pads in PCB)


Power, Clock, Config, Data transmitted

Detailed description can be found <u>JINST14 P07010</u>




GigaTracKer in operation

- Fully operational since September 2016 (first station installed in 2014)
- Few noisy/dead pixels (< 1% per station)
- Nominal bias voltage 100V
- Beam intensity was increasing over the years of data taking, being 35% of nominal in 2016 and reaching 100% in 2018.





Kinematic performance



Sample of $K^+ \rightarrow \pi^+\pi^+\pi^-$ is used to align GTK and STRAW spectrometer and equalize measurement of momentum


Physics performance from a sample of $K^+ \rightarrow \pi^+ \pi^0$ matches design performance (16 µrad)




Precise time measurement with GTK

Offline calibration is necessary to correct for time walk and delays in the pixel matrix:

- Pixel time offsets extracted at MPV of ToT for each pixel (~3ns variation along a column)
- Time walk correction computed per chip (~300 ps/1ns ToT)
- Reference time: KTAG (70 ps resolution)
- Calibration is done once for each group of ~1500 spills



GTK performance in 2016-2017



Conditions:


- n-in-p sensors @100 V Measurement methods:
- Time difference between GTKs KTAG RICH (σ_{t} < 100 ps)
- Time difference between the 3 GTK stations

Observations:

- Hit time resolution: 130 ps (best 115 ps @250 V)
- Track time resolution: 75 ps (best 65 ps @250 V)
- ~20% time resolution degradation observed by the end of 2017, which is still better than the design requirements
- Peak(average) fluence by the end of 2017: $1.3(0.26) \times 10^{14}$ 1 MeV n_{eq} /cm²
- Overall efficiency: 97%, 99% per station



GTK performance in 2018-2022



Conditions:


- 15 modules in total were used as GTK1,2, 3 and 0 from 2016 to 2023, a few modules were used for several years
- All modules, except one are n-in-p
- Nominal bias volage (HV) is 100 V
- After the end of each data taking period, modules are removed and stored in cold

Observations:

- Hit time resolution: 120 ps best, 260 ps worst(p-in-n @ 100V)
- ~30% time resolution degradation for the most irradiated module by the end of 2018:
 - Starting from 2021 bias voltage adjustment was used to improve performance of old modules and keep resolution stable



GTK performance in 2018-2022



- The most irradiated module by 2023 with a peak (average) fluence of 2.5(0.81) × 10^{14} 1MeV n_{eq} /cm² provided 135 ps time resolution @250 V at the end of 2022 data taking
- Resolution of p-in-n sensors is factor of 2 of n-in-p and scales differently with fluence



Prospects

HIKE (High-Intensity Kaon experiments): a long-term rare kaon decay programme @ CERN SPS (>~2030)


• Multiple phases, K^+ and K_L decay experiments


HIKE Phase 1: $K^+ \rightarrow \pi^+ \nu \bar{\nu}$ at 5% precision in 4 years:

• 4x intensity on GTK → new challenging specifications

E field

- Possible synergies with ongoing developments:
 - TimeSPOT (3D sensors)







Beam rate	>3GHz
Peak particle flux	>8.0 MHz/mm ²
Peak radiation	2.35×10 ¹⁵ 1MeV n _{eq} /cm²/200 days
Efficiency	99%
Pixel size	<300 µm
Pixel time resolution	<50ps
Material budget	0.3-0.5%X ₀
Detector size	60.8x27mm ²

LOI: <u>arXiv:2211.16586</u>

Summary

- The GigaTracKer is the NA62 4D beam tracker and is essential for the precise measurement of BR($K^+ \rightarrow \pi^+ \bar{\nu} \nu$)
- The detector is fully operational since 2016 and successfully taking data
- The single hit time stamp resolution of 115 ps was achieved, surpassing the design resolution.
- Analysis of data collected in 2018-2021 demonstrated that GTK can maintain 135 ps hit time stamp resolution even after being irradiated up to 2.5(0.81) × 10^{14} 1MeV n_{eq} /cm², corresponding to ~120 days at nominal intensity.


Prospects:

- Analysis of data is ongoing + preparing to restart data taking again next year
- New experimental challenges ahead: GTKx4 for HIKE experiment @ CERN SPS





16/10/2023



Backup slides


16/10/2023

Operation and performance of the NA62 GigaTracKer

21



Time resolution of p-in-n and n-in-p





Time resolution of p-in-n and n-in-p





Figure 13. Standard deviation for the centroid time for sensor thickness of a) $d = 200 \,\mu\text{m}$ and b) $d = 50 \,\mu\text{m}$ and $V = 200 \,\text{V}$, assuming uniform charge deposit and a square readout pad. The horizontal line represents centroid time resolution from eq. (4.5) due to Landau fluctuations only. The two curves in the plots represent the effect of weighting field fluctuations where either the electrons or the holes move towards the readout pad.

2017 JINST 12 P11017


16/10/2023



Time resolution+temperature




16/10/2023



16/10/2023



TDCPix ASIC



Two sections:

- 40x45 pixel matrix: preamplifier, discriminator (5 bit DAC), asynchronous transmission lines
- end-of-column:TDCs, digital logic
- Organized by columns (1 column=45 pixels)
 - group of 5 pixels send signals to 1 Hit Arbiter
 - 1 Hit Arbiter sends signal to a dual TDC (leading and trailing edge): 9 TDCs per column/360 TDCs per chip

Per column pair:

• 1 DLL provides the fine timestamp, 97 ps bin of the TDC

Per groups of 10 columns: 3.2 Gb/s data transmitter

Power consumption: 3.5W per chip (>90% is in EOC region)