

Positron-Range Correction for an On-Chip PET Scanner using Deep Learning

Monday, 20 May 2024 18:00 (5 minutes)

Background: Organs-on-Chips (OOCs) are a novel technology that aim to mimic the functions and physiology of human organs in a laboratory setting. Positron Emission Tomography (PET) is a widely-used imaging modality that enables non-invasive monitoring of biological processes in vivo. However, the spatial resolution of current small-scale PET systems is not sufficient for OOC imaging. One of the main factors limiting the spatial resolution of a PET scanner is the positron range, which is the distance that a positron travels before it collides with an electron. In this study, we present a novel Deep Learning (DL)-based approach for correcting the positron-range effect in our previously introduced On-Chip PET scanner.

Results: We created a dataset of pairs of non-corrected and corrected images using a Monte-Carlo simulation of a realistic OOC phantom and a fully three-dimensional Maximum-Likelihood Expectation-Maximization (MLEM) iterative reconstruction algorithm. Our results demonstrate the effectiveness of the DL-based positron-range correction algorithm in improving the overall quality of the reconstructed images.

Field

Software and quantification

Primary author: CLEMENT, Christoph (Department of Nuclear Medicine, Inselspital, University Hospital Bern, Bern, Switzerland)

Co-authors: PAGANO, Fiammetta (EP Department, CERN, Geneva, Switzerland; Physics Department, University of, Milano-Bicocca, Milan, Italy); PIZZICHEMI, Marco (EP Department, CERN, Geneva, Switzerland; Physics Department, University of, Milano-Bicocca, Milan, Italy); TERRAGNI, Giulia (EP Department, CERN, Geneva, Switzerland; Physics Department, Technical University of Vienna, Vienna, Austria); KRUTHOF-DE JULIO, Marianna (Department for Biomedical Research, Inselspital, University Hospital Bern, Bern, Switzerland); ZIEGLER, Sibylle (Department of Nuclear Medicine, LMU Klinikum, University Hospital, Munich, Munich, Germany); ROMINGER, Axel (Department of Nuclear Medicine, Inselspital, University Hospital Bern, Bern, Switzerland); AUFRAY, Etienne (EP Department, CERN, Geneva, Switzerland); SHI, Kuangyu (Department of Nuclear Medicine, Inselspital, University Hospital Bern, Bern, Switzerland)

Presenter: CLEMENT, Christoph (Department of Nuclear Medicine, Inselspital, University Hospital Bern, Bern, Switzerland)

Session Classification: Poster Session

Track Classification: High-performance preclinical and organ-specific systems