

Validation of CT-free Template-Based Attenuation Correction in Brain PET Imaging

Monday, 20 May 2024 18:45 (5 minutes)

This study aims to validate a template-based attenuation correction (AC) method for clinical brain PET systems without CT, previously proposed for the new NeuroLF brain-PET. Method: The validation process involved simulating a diverse (patients/tracers) set of clinical images. This is done by transforming 142 reference PET/MR and 8 PET/CT brain images to NeuroLF image space, converting them to sinograms, then uncorrecting the sinograms for attenuation and scatter using original patient-derived attenuation maps. These uncorrected sinograms serve as a surrogate for real clinical raw data. From this, images were then re-reconstructed with both patient-specific reference and tracer-specific template attenuation maps and compared using a range of similarity metrics. Results: The stability of the template-based AC method across a diverse set of tracers for brain PET imaging could be shown. The used metrics identified cases where the MR-to-CT conversion was inadequate, but were less capable of detecting failures intentionally induced in the template map co-registration. Tracer-specific templates showed better performance than a universal (i.e. [18F]-FDG derived) template. Conclusion: The validation confirms the adequacy of the template-based AC method for clinical routine brain-PET imaging, with a recommendation for visual inspection to detect co-registration failures. We aim to further validate the template-based attenuation maps in a blinded clinical evaluation.

Field

Software and quantification

Primary author: MIKHAYLOVA, Ekaterina (Positrono AG, Zurich, 8005 Switzerland)

Co-authors: JEHL, Markus (Positrono AG, Zurich, 8005 Switzerland); THIELEMANS, Kris (University College London); Dr SACCO, Ilaria (Positrono AG, Zurich, 8005 Switzerland); Dr AHNEN, Max (Positrono AG, Zurich, 8005 Switzerland); Dr SATTLER, Bernhard (University Hospital Leipzig, Department of Nuclear Medicine, Germany); Prof. BARTHEL, Henryk (University Hospital Leipzig, Department of Nuclear Medicine, Germany); Prof. SABRI, Osama (University Hospital Leipzig, Department of Nuclear Medicine, Germany); FISCHER, Jannis (Positrono AG, Zurich, 8005 Switzerland)

Presenter: JEHL, Markus (Positrono AG, Zurich, 8005 Switzerland)

Session Classification: Poster Session

Track Classification: High-performance preclinical and organ-specific systems