PSMR2024 10th Conference on PET, SPECT, and MR Multimodal Technologies, Total Body and Fast Timing in Medical Imaging

Contribution ID: 86

Type: Poster

## Estimating skeletal muscle metabolism from 31P MRSI at 7 Tesla using a dual-tuned volume coil preliminary in-vivo results

Monday, 20 May 2024 18:55 (5 minutes)

Phosphorus Magnetic Resonance Spectroscopy (31P MRS) is a non-invasive imaging technique that estimates concentrations of high-energy phosphates, offering insights into cell metabolism and neuromuscular disorders. Its clinical adoption is hindered by the need for specialized coils, but using a volume coil dual-tuned to 1H and 31P for 3D Magnetic Resonance Spectroscopy Imaging (3D-MRSI) can overcome this difficulty. Our approach allows for fast acquisition with improved spectral resolution at Ultra-High Field (7T). This approach can potentially enhance 31P metabolic imaging in clinical settings.

## Field

Software and quantification

**Primary authors:** BUCHIGNANI, Bianca (IRCCS Stella Maris); ARINGHIERI, Giacomo (IMAGO7 Foundation); BIAGI, Laura (IRCCS Stella Maris); TOSETTI, Michela (IRCCS Stella Maris and IMAGO7 Research Foundation); CECCHI, Paolo (IMAGO7 Foundation); AMBROSI, Pierfrancesco (IRCCS Stella Maris); BATTINI, Roberta (Università di Pisa); SCHULTE, Rolf F. (GE HealthCare)

**Presenter:** BIAGI, Laura (IRCCS Stella Maris)

Session Classification: Poster Session

Track Classification: Next-gen MR