

# Rigid Motion Detection for Abrupt Motion in FDG Brain PET Imaging

*Monday, 20 May 2024 19:00 (5 minutes)*

In this study, we employed a dedicated brain positron emission tomography (PET) scanner to simulate controlled motion through the use of a voxelised phantom. The resulting output was converted to listmode and subsequently transformed into pseudo-displacement time series via a moving mean technique. To locate time points of motion, we applied a statistical approach known as bottom-up segmentation, a form of change point detection. This facilitated the identification and separation of frames into motion-free segments. Bottom-Up Segmentation consistently detected spatial resolutions as fine as 2mm and identified motions of 3mm within intervals, up to, 6 seconds consistently.

## Field

Software and quantification

**Primary author:** DAO, Viet (University of Leeds)

**Co-authors:** TSOUNPAS, Charalampos (University Medical Centre Groningen); MIKHAYLOVA, Ekaterina (Positron AG)

**Presenter:** DAO, Viet (University of Leeds)

**Session Classification:** Poster Session

**Track Classification:** PET/MR and SPECT/MR