PSMR2024 10th Conference on PET, SPECT, and MR Multimodal Technologies, Total Body and Fast Timing in Medical Imaging

Contribution ID: 53

Type: Poster

Quality control of plastic scintillators for the total-body J-PET scanner

Wednesday, 22 May 2024 16:10 (5 minutes)

Total-body Jagiellonian positron emission tomography (TB-J-PET) is based on long plastic scintillators [1] which decrease cost of the scanner [2]. TB PET scanners enable positronium imaging [3], measurements of polarization of photons [4] and beam therapy monitoring [5]. Development of TB-J-PET requires application of transparent plastic scintillators with low light attenuation [6] to build modules with SiPMs attached at both ends of the scintillators. For modular TB-J-PET construction we choose BC-408, one of the most transparent plastic scintillator from our previous measurements [7]. Purpose of this research is to verify quality of received scintillators. Transmittance at the wavelength of maximum emission through 6 mm thick scintillator and technical attenuation length along 330 mm long scintillator were measured on linear CCD array spectrometer for random selected scintillators from each delivered batch. Additionally, scintillators optical homogeneity was measured on light transfer setup consisting of exciting LED and photodiode matrix.

- [1] S. Vandenberghe, EJNMMI Physics 7 (2020) 35
- [2] P. Moskal et al., Physics in Medicine and Biology 66 (2021) 175015
- [3] P. Moskal et al., Science Advances 7(42) (2021) eabh4394
- [4] P. Moskal et al., Nature Communications 15 (2024) 78
- [5] K. Parodi, Journal of Medical Physics 33 (2023) 22
- [6] Ł. Kapłon, IEEE Transactions On Nuclear Science 67(10) (2020) 2286
- [7] Ł. Kapłon et al., Nucl. Instrum. Methods Phys. Res. A 1051 (2023) 168186

Field

Detectors and electronics

Primary author: KAPŁON, Łukasz (Jagiellonian University)

Presenter: KAPŁON, Łukasz (Jagiellonian University)

Session Classification: Poster Session

Track Classification: Total body imaging