





Weekly shift summary

02-09/08/2023

Dayron Ramos - Politecnico and INFN Bari

Summary

- Not special issues during the week 02-09/08/2023
- Weekly scan @ source off 02/08/2023 ScanID 622 (not included SHiP), stability scan 172 (all chamber)
- SHiP chamber had a discharge problem at the beginning of the shift and was shutting down all the HV module -> the problem was solved grounding the chamber
- ALTAS chamber with strong current fluctuations (contact Barbara/Marco)

