

RPC eco-gas at GIF++ Weekly report

Barbara Liberti and Marco Sessa November 30th, 2023



Istituto Nazionale di Fisica Nucleare

Stability run 194


- Stability run 194 started on Thursday November 23rd around 9:00 and stopped on Wednesday November 29th around 10:30
- All chambers included, except ATLAS (temporarily disconnected from the setup) and ALICE

• Major problems during the shift:

- Problems with the WET MFC regulation —> no response —> Luca changed the WET to 100% and regulated the DRY (as it was in the past) —> now it works —> e-log <u>https://rpcecogas.web.cern.ch/rpcecogaselog/440</u>
- Data logger crashed many times (order of 10 times this week) —> also during the second HV scan with source OFF (scan no visible in grafana) and sometimes without alarm e-mail notification

• Major achievement of the week:

Luca and Gianluca wrote a script to extract the humidity value and log it to the ecogas dashboard —>
e-log <u>https://rpcecogas.web.cern.ch/rpcecogaselog/441</u>







Stability run 194



Stability run 194






HV scans with source OFF

HVscan run registry


Scan ID	Scan type	Start time	
	Select scan type		
000703 🖻	Current scan	2023-11-29 16:09	
000702 🖻	Current scan	2023-11-29 11:07	

Wrong label -> we hope this is not a problem






HV scan 702




Similar results for scan 702 and 703 -> only a 30 uA jump observed for SHIP in scan 702 around 8000 V

HV scan 703




Similar results for scan 702 and 703 -> only a 30 uA jump observed for SHIP in scan 702 around 8000 V

HV scan 701 from last week (for comparison)



Currents seem to be comparable w.r.t. last week scan (701)

Residual HFO today at 12:30



Next stability run

- 12:30)
- For this reason the run 195 was stopped this morning —> maybe it can be also deleted
- Next stability run has to start when the source will be switched ON

Instructions for next shifter

- ATLAS and ALICE still excluded
- EPDT, RE1-1, SHiP, KODEL-H to be included
- (only during scan!)
- For stability scan use template 194

Chamber/gap	Enabled	Working voltage (V)	Standby voltage (V)	i0 (uA)
EPDT-RPC25		9000	6500	20
SHiP-1-6		8800	5000	200
RE1_1_001-BOT		10600	6500	200
RE1_1_001-TN		10600	6500	200
RE1_1_001-TW		10600	6500	200
KODEL-H-BOT		7700	5000	50
KODEL-H-TOP		7700	5000	50

From Giuliana's slides of last week - instructions still valid

• We started a stability run yesterday evening at 19:26 (Run 195) —> but the source was off during the night and it is still OFF now (at

• For current scan use template 701 -> EPDT change Itrip to 100uA from telnet

