Channeling 2024



Contribution ID: 42

Type: invited

Implementing Capillary Design for Reliable VHEE Beam Delivery

Thursday, 12 September 2024 13:05 (30 minutes)

Very High Energy Electron (VHEE) radiotherapy is gaining attention for its potential to revolutionize cancer treatment [1]. VHEE employs high-energy electrons (~250 MeV) accelerated to extremely high speeds, which are precisely targeted at deep-seated tumors. This method offers significant penetration and optimal dose distribution, effectively targeting tumors while sparing healthy tissues. The precision of VHEE supports high dose-rate irradiation, promising improved treatment outcomes and fewer side effects. However, the widespread adoption of VHEE is currently limited by the availability of hospital-scale accelerators, a challenge being addressed by advancements in high-gradient laser-plasma accelerators (LPAs) [2].

Our proposed method focuses on accelerating electrons using lasers in capillary discharge [3] at I-LUCE (INFN-Laser indUCEd radiation production), a new laser facility in Catania, Italy. We are designing a capillary system there for generating electrons optimized for dual FLASH/VHEE modes. Different capillary geometries (length, diameter, channel shape) will be studied to enhance high-quality electron beam production.

Capillaries demonstrate exceptional capability in producing high-quality electron beams with precise energy control, despite requiring meticulous laser focusing and a complex setup. They effectively guide and confine laser pulses and plasmas over extended distances, facilitating higher acceleration gradients and superior control over beam emittance. The use of capillaries for e-LPAs holds particular promise for VHEE applications within the Laser Wakefield Acceleration (LWFA) scheme, enabling energies and fluxes reaching several GeV to be achieved [4]. Therefore, by leveraging the I-LUCE laser's characteristics—50-350 TW power, 1-7 J energy, 23-150 fs pulse duration, 10^17-10^19 W/cm^2 intensity, and 1-10 Hz repetition rate—along with a plasma density of 10^17-10^19 cm^(-3), we can produce electron beams with energies ranging from 0.1 to 3 GeV. These beams will have a charge of 30-400 pC and contain 10^8-10^9particles per pulse.

References

- 1. T Fuchs et al, Phys. Med. Biol. 54 (2009) 3315.
- 2. T Tajima and J M Dawson, Phys. Rev. Lett. 43 (1979) 267.
- 3. S Arjmand et al, JINST. 18 (2023) P08003.
- 4. W P Leemans et al, Phys. Rev. Lett. 113 (2014).

Primary author: Dr ARJMAND, Sahar (INFN-LNS)

Co-authors: Dr SCIUTO, Alberto (INFN-LNS); Dr PAPPALARDO, Alfio Domenico (INFN-LNS); Mr HASSAN, Ali (INFN-LNS); Ms KURMANOVA, Alma (INFN-LNS); Mr AMATO, Antonio (INFN-LNS); Mr MANNA, Carmelo (INFN-LNS); Mr OLIVA, Demetrio (INFN-LNS); Dr ABUBAKER, Farmesk (INFN-LNS); Dr FAROKHI, Fateme (INFN-LNS); Mr VINCIGUERRA, Francesco (INFN-LNS); Dr CIRRONE, GA Pablo (INFN-LNS); Dr CUTTONE, Giacomo (INFN-LNS); Dr PETRINGA, Giada (INFN-LNS); Mr CANTONE, Giovanni (INFN-LNS); Dr SUAREZ, Jose (INFN-LNS); Dr GUARRERA, Mariacristina (INFN-LNS); Dr CATALANO, Roberto (INFN-LNS); Dr FATTORI, Serena (INFN-LNS)

Presenter: Dr ARJMAND, Sahar (INFN-LNS)

Session Classification: Laser/Plasma & Channeling