

Krypton in XENON


Sebastian Lindemann

Max-Planck-Institut für Kernphysik
Heidelberg

Low Radioactivity Techniques, 2013

XENON100

Dark Matter Search




E. Aprile et al. (XENON100), Phys. Rev. Lett. 109, 181301 (2012), arXiv:1207.5988.

XENON100: Effect of ^{85}Kr on Background Level

Krypton Distillation




+



E. Aprile et al. (XENON100), Phys. Rev. Lett. 107, 131302 (2011)

=



"100 Live Days"
48kg FV

22 mDRU

"225 Live Days"
34kg FV

5.3 mDRU

similar to:

K. Abe et al. (XMASS Collaboration),
Astropart. Phys. 31 290 (2009)

E. Aprile et al. (XENON100), Phys. Rev. Lett. 109, 181301 (2012)


The Need for an Off-Line Analysis Tool

^{85}Kr limits search for Dark Matter

- ⇒ need to remove ^{85}Kr
 - krypton distillation column
- ⇒ need (off-line) analysis tools
 - delayed coincidence
 - spectral fit

Too little ^{85}Kr (for analysis)

- 1 ppt Kr in Xe:
- about 3200 ^{85}Kr atoms in 35 kg LXe
 - about 0.6 ^{85}Kr -decays/day/35 kg LXe
 - ⇒ months of data needed
- ⇒ make use of known (see next slide) $^{85}\text{Kr}/^{nat}\text{Kr}$ ratio



Off-line krypton assay

RGMS (Rare Gas Mass Spectrometer)

Insertion: The $^{85}\text{Kr}/\text{nat Kr}$ Ratio

Geophysical Literature

$$^{85}\text{Kr}/\text{nat Kr} \approx 2 \times 10^{-11} \text{ mol/mol}$$


(X. Du et al., Geophysical Research Letters, 30(20), Oct 2003)

LNGS air sample using miniaturized proportional counters

^{85}Kr in Hall A air sample:

$$\begin{aligned} & 1.3 \pm 0.2 \text{ mBq/liter (air sample drawn on 1 Oct 2009; S. Lindemann diploma thesis, Heidelberg 2009)} \\ & \Rightarrow 2.1 \pm 0.3 \times 10^{-11} \text{ mol/mol} \end{aligned}$$


Monitoring in (South-)Germany by BfS



J. Bieringer et al. (Appl. Rad. Isot. 67)


$$\Rightarrow {}^{85}\text{Kr}/\text{nat Kr} = 2 - 4 \times 10^{-11} \text{ mol/mol}$$

Introduction to the RGMS




- Only inert gases (RGMS = Rare Gas Mass Spectrometer)
- Pressure $< 10^{-6}$ mbar
- Below 10^{-12} cm³ (STP) krypton detectable

Introduction to the RGMS



- Only inert gases (RGMS = Rare Gas Mass Spectrometer)
- Pressure $< 10^{-6}$ mbar
- Below 10^{-12} cm³ (STP) krypton detectable
- 1ppt krypton $\Rightarrow 1$ cm³ (STP) xenon needed \Rightarrow pressure $\gg 10^{-6}$ mbar ↴


Introduction to the prepended Cryogenic Gas Chromatograph (pCGC)



- He matrix gas is cleaned
- Kr stays less time in adsorber than Xe
- Kr atoms are stored; Xe bulk is rejected
- Kr is transferred to MS


Krypton-Xenon separation

- Cryogenic gas chromatography at -80°C
- Cutting around 15 min
⇒ effective rejection of xenon bulk
- Amount of He matrix gas:
 $\sim 10 - 100 \times$ size of Xe sample



RGMS: Procedure Blank and Detection Limit

Detection Limit (DL)




- Procedure Blank limits absolute amount of ^{nat}Kr detectable
- Maximal size of xenon sample determines RGMS performance

$$\begin{aligned} \text{DL} &= (^{nat}\text{Kr})_{min}/(\text{Xe})_{max} \\ &\approx 1.3 \times 10^{-12} / 2 \text{ ccSTP/ccSTP} \\ &= 0.7 \text{ ppt (90\% C.L.)} \end{aligned}$$

Procedure Blank

- Procedure Blank: 1×10^{-12} ccSTP ^{nat}Kr
- ^{nat}Kr in He matrix gas < 0.1 ppt



Estimation of RGMS Errors

Errors systematically affecting all measurements

- Absolute calibration of standard (eg1&eg2)
- Xenon sample size determination
 - Size of calibration volume
 - Accuracy of pressure gauge
- Krypton acceptance of chromatographic separation (pCGC)

Errors affecting individual measurements

- Instability of electronics (ion-source, ion-multiplier, ...) wrt preceding calibration
- Variations in krypton acceptance of pCGC due to manual operation

Total error of RGMS about 20%


Drawing Xenon Samples for RGMS Measurements



Available sampling containers

- *Xe-pipettes*: 5 - 20 ccSTP samples enclosed in stainless steel, fully metal sealed volumes
- *Xe-bombs*: 1 liter, fully metal sealed (CF16), stainless steel vessels
- Or: connecting xenon gas cylinder directly to RGMS

Conversion ppt → mDRU



see also
M. Selvi's talk
on Friday

E. Aprile et al. (XENON100), Phys. Rev. D 83, 082001 (2011)

DRU (Differential Rate Unit) \equiv cts/(keV_{ee} · kg · day)

Using $^{85}\text{Kr}/^{nat}\text{Kr} = 2 \times 10^{-11}$ mol/mol, the conversion to mDRU for XENON100 reads:

$$100 \text{ ppt} \hat{=} 3.9 \text{ mDRU}$$

Results from RGMS Measurements

XENON100: 100 Live Days publication

Background fit: 700 ± 100 ppt ($^{85}\text{Kr}/\text{nat Kr} = 10^{-11}$ mol/mol)

RGMS: 360 ± 70 ppt

$\Rightarrow 14$ mDRU from ^{85}Kr

XENON100: 225 Live Days publication

Delayed Coincidence analysis: 18 ± 8 ppt

RGMS: 14 ± 3 ppt

$\Rightarrow 0.6$ mDRU from ^{85}Kr

Performance of XENON100 Distillation Column

RGMS: reduction factor = 350 ± 100 (single pass starting from 2.8 ± 0.6 ppb)

Right now in the XENON100 detector

Less than 1.3 ppt (90% C.L.)




\Rightarrow less than 0.05 mDRU from ^{85}Kr

Conclusions


- ^{85}Kr is a potentially dangerous source of background in XENON100 (and XENON1T).
- The RGMS, an off-line krypton in xenon analysis tool, is fully operational with a detection limit below 1 ppt.
- The XENON100 distillation column has proven its capability to degrade krypton to a sub-dominant background.
- XENON100 on-line data-analysis techniques (delayed coincidences and spectral fit) and off-line krypton assay show excellent agreement.
- The achieved krypton level of $< 1.3 \text{ ppt}$ in XENON100 is an important step towards the requirement of $< 0.1 \text{ ppt}$ needed for XENON1T.

Backup Slides

XENON100: 3-dimensional Position Reconstruction








XENON100: 3-dimensional Position Reconstruction




XENON100: Nuclear and Electronic Recoils

nuclear recoil

electronic recoil



RGMS Calibration



Daily calibration: eg1&eg2

- Permanently installed large gas reservoir
- Pipette off a fixed fraction ($\sim 10^{-5}$) of large gas reservoir
 - 2.2 ± 0.4 ccSTP ^{84}Kr
 - 6.7 ± 1.1 ccSTP ^{86}Kr
- Use before (and after) a measurement to calibrate ion-multiplier response

Absolute calibration of eg1&eg2

Commercial calibration gas (helium) with 10 ppm ^{nat}Kr (2% error)

- extract small amount of gas (~ 0.01 ccSTP): 14% error
- dilute calibration gas (by factor 25 - 200): 2.5% error
- freeze krypton to cold finger, remove helium by pumping: < 2% error

