

## Proton Computed Tomography system: recent results and upgrade status

M. Bruzzi<sup>1,2</sup>, M. Bucciolini<sup>2,3</sup>, M. Carpinelli<sup>4,5</sup>, C. Civinini<sup>2</sup>, G.A.P. Cirrone<sup>6</sup>, G. Cuttone<sup>6</sup>, D. Lo Presti<sup>7,8</sup>, S. Pallotta<sup>2,3</sup>, C. Pugliatti<sup>7,8</sup>, N. Randazzo<sup>7</sup>, F. Romano<sup>6,9</sup>, M. Scaringella<sup>10</sup>, V. Sipala<sup>4,5</sup>, C. Stancampiano<sup>7,8</sup>, C. Talamonti<sup>2,3</sup>, E. Vanzi<sup>11</sup>, M. Zani<sup>2,3</sup>

<sup>1</sup> Dipartimento di Fisica e Astronomia, Università di Firenze, Via G. Sansone, Sesto Fiorentino, Firenze, Italy

<sup>2</sup> INFN Sez. Firenze, Via G. Sansone 1, Sesto Fiorentino, Firenze, Italy

<sup>3</sup> Dipartimento di Scienze Biomediche, Sperimentali e Cliniche, Università di Firenze, Firenze, Italy

<sup>4</sup> INFN sez. di Cagliari, S.P. Sestu, Monserrato, Cagliari, Italy

<sup>5</sup> Dipartimento di Chimica e Farmacia, Università di Sassari, Via Vienna 2, Sassari, Italy

<sup>6</sup> INFN Laboratori Nazionali del Sud, Via. S. Sofia 62, Catania, Italy

<sup>7</sup> INFN sez. Catania, Via. S. Sofia 64, Catania, Italy

<sup>8</sup> Dipartimento di Fisica e Astronomia, Università of Catania, Via. S. Sofia 64, Catania, Italy

<sup>9</sup> Museo Storico della Fisica e Centro Studi e Ricerche "Enrico Fermi", Piazza del Viminale 1, Rome, Italy

<sup>10</sup> Dipartimento di Ingegneria Industriale, Università di Firenze, Via S. Marta 3, Florence, Italy

<sup>11</sup> SOD Fisica Medica, Azienda Ospedaliero Universitaria Careggi, Largo Brambilla 3, Florence, Ital

Contact email: [mara.bruzzi@unifi.it](mailto:mara.bruzzi@unifi.it).

Proton Computed Tomography (pCT) is a medical imaging technique based on the use of proton beams with energies above 200MeV to directly measure stopping power distributions inside the tissue volume. Prima (PRoton IMAGING) is an Italian collaboration working on the development of a pCT scanner based on a tracker and a calorimeter to measure single protons trajectory and residual energy. The tracker is composed by four planes of silicon microstrip detectors to measure entry and exit positions and angles. Residual energy is measured by a calorimeter composed by YAG:Ce scintillating crystals. A first prototype of pCT scanner, with an active area of about 5x5 cm<sup>2</sup>, has been constructed and characterized with 60 MeV protons at the INFN Laboratori Nazionali del Sud Catania (Italy) and with 180 MeV protons at Svedberg Laboratory Uppsala (Sweden). A new pre-clinical prototype with an extended active area up to 5x20 cm<sup>2</sup>, real time data acquisition and a data rate up to 1 MHz is under development. A description of the two prototypes will be presented together with first results concerning tomographic image reconstruction.