



Contribution ID: 53

Type: Poster

P21 - A compact gas ionisation direct-STIM detector for MeV ion microscopy

Friday, 11 July 2014 13:00 (1 hour)

Scanning Transmission Ion Microscopy is a powerful technique that yields structural information in sub-cellular whole cell imaging [1,2]. Usually, a Si p-i-n diode is normally used in direct-STIM measurements as a detector. However, this is sensitive to radiation damage because when used to image deep sub- μm areas the ion fluence is high even for moderate numbers of ions per pixel. This leads to a shift in charge carrier collection efficiency which appears as an apparent degradation in energy resolution [3,4]. Gas ionisation detectors are intrinsically insensitive to radiation damage. Therefore, a compact gas ionization direct-STIM detector [4] is based on the Geiger-Muller geometry without a Frisch-grid is being developed for use in a MeV ion microscope based on a standard Oxford triplet lens and scanning system. In this work, the critical issues in designing the detector are a large the entrance window area to accept a scanned beam and obscuring as small a solid angle as possible to facilitate installation of proton induced fluorescence detector [2] while maintaining a small capacitance (10.2 pF) to achieve low noise. Tests with isobutane against a calibrated capacitor show that a resolution of about ~ 20 keV fwhm could be achieved which allowed imaging of Ren cells on a Si₃N₄ support and an effective ionisation energy of 21 eV at pressures above 600 mbar and bias voltage >600 V. The resolution is most probably limited by the charge sensitive amplifier resolution which is estimated to be ~ 180 e- rms in the current configuration.

References

- [1] F. Watt, X. Chen, A.B.D. Vera, C. N. Udalgama, R. Minqin, J. A. van Kan, A.A. Bettiol, The Singapore high resolution single cell imaging facility, Nucl. Instr. and Meth. B 269 (2011) 2168.
- [2] R. Norarat, V. Marjomäki, X. Chen, M. Zhaohong, R. Minqin, C. Chen, A.A. Bettiol, H.J. Whitlow and F. Watt, Ion-induced fluorescence imaging of endosomes, Nucl. Instr. and Meth. B 306 (2013) 113.
- [3] R. Minqin, J.A. van Kan, A. A. Bettiol, L. Daina, C.Y. Gek, B.B. Huat, H.J. Whitlow, T. Osipowicz, and F. Watt, Nano-imaging of single cells and tissue using MeV ions, Nucl. Instr. and Meth. B 260 (2007) 124.
- [4] A.M. Müller, A. Cassimi, M. Döbeli, M. Mallepell, I. Monnet, M.J. Simon, M. Suter, and H.-A. Synal, A new mini gas ionization chamber for IBA applications, Nucl. Instr. and Meth. B 269 (2011) 3037.

Primary author: Dr NORARAT, Rattanaporn (Faculty of Science and Agriculture, Rajamangala University of Technology Lanna Chiang Rai, 57120 Chiang Rai, Thailand)

Co-authors: Mr GUIBERT, Edouard (University of Applied Sciences, Haute Ecole Arc Ingénierie, (HES-SO) Eplatures-Gris 17, CH-2300 La Chaux-de-Fonds, Switzerland); Prof. WHITLOW, Harry James (University of Applied Sciences, Haute Ecole Arc Ingénierie, (HES-SO) Eplatures-Gris 17, CH-2300 La Chaux-de-Fonds, Switzerland); Dr STOPPINI, Luc (inSTI, HEPIA, Rue de la Prairie 4, CH-1202 Genève, Switzerland.); Dr JEANNERET, Patrick (University of Applied Sciences, Haute Ecole Arc Ingénierie, (HES-SO) Eplatures-Gris 17, CH-2300 La Chaux-de-Fonds, Switzerland)

Presenter: Dr NORARAT, Rattanaporn (Faculty of Science and Agriculture, Rajamangala University of Technology Lanna Chiang Rai, 57120 Chiang Rai, Thailand)

Session Classification: Poster Session with Cheese and Wine