

# Compatibility of Metal Additive Manufactured Tungsten Collimator for SPECT/MRI Integration

Wednesday, May 20, 2015 11:00 AM (1h 30m)

We optimized the MR-compatibility of a novel tungsten collimator, produced with metal additive manufacturing that is part of a microSPECT insert for a preclinical SPECT/MRI scanner. We characterized the current density due to the gradient field and adapted the collimators by smart design to reduce the induced eddy currents.

The z-gradient coil and the collimator were modeled with SEMCAD. The gradient strength was 510 mT/m, the gradient efficiency was about 3.4 mT/m/A. The setup was simulated with a working frequency of 10 kHz. The system consists of 7 identical collimators and digital silicon photomultipliers assembled in a ring. We evaluated the global reduction in current density  $J(\text{reduction})$  based on the sum of all current densities in the collimator. We applied the following optimizations on the collimator:

1. We reduced the excessive material in the flanges.
2. We applied horizontal slits of 2 mm in the collimator surface.
3. We reduced material in the core; the photons are attenuated before they reach the core. The collimator will need a supporting structure.
4. The supporting structure can be avoided by using two vertical slits in the middle of the collimator.
5. We used a Z-shaped slit instead of the vertical slit.

Results of simulations show that smaller flanges reduce the current density with 23%. The horizontal slits reduce the eddy currents with 6%. Using less material in the core or applying vertical slits results in the same reduction of current density. However, the vertical slits are cheaper because a hollow collimator requires supporting structures during production. Both can be combined if z-shaped slits are used to prevent attenuation problems. The reduction is then 27%. Finally, when all previous adaptations are combined, the reduction in eddy currents is about 56.3%.

**Primary author:** SAMOUDI, Amine (Ghent University/iMinds)

**Co-authors:** Dr VERMEEREN, Günter (INTEC, Ghent University/iMinds, Ghent, Belgium); Ms VAN AUDENHAEGE, Karen (MEDISIP, IBBT-IBiTech, Ghent University); Prof. MARTENS, Luc (INTEC, Ghent University/iMinds, Ghent, Belgium); Prof. VAN HOLEN, Roel (Ghent University); Prof. JOSEPH, Wout (INTEC, Ghent University/iMinds, Ghent, Belgium)

**Presenter:** SAMOUDI, Amine (Ghent University/iMinds)

**Session Classification:** Session 11 - Poster Session II

**Track Classification:** 4 - Advances in MR-SPECT instrumentation