Type: Talk

PET/MR: Improvement of the UTE μ-Maps Using Modified MLAA

Tuesday, 19 May 2015 11:00 (15 minutes)

For a quantitative analysis in positron emission tomography (PET) or single-photon emission computed tomography (SPECT), attenuation correction (AC) is mandatory. CT-scans or transmission scans are common tools for determination of the attenuation μ -map, but in the case of a PET/MR hybrid system it is difficult to associate one of these scans. Many techniques have been developed in order to improve AC for PET/MR. Some methods are based on template- or atlas techniques, other methods apply a segmentation technique based on Dixon or UTE (Ultrashort Echo Time) MR to create the µ-map, followed by a standard OSEM reconstruction (OSEM/DIXON and OSEM/UTE). A different approach for AC has been developed by employing the emission sinogram data in the µ-map derivation. In this context, we modified the iterative MLAA (Maximum-Likelihood reconstruction of Attenuation and Activity) algorithm to improve the resulting emission image from the PET/MR system. We constrained the attenuation map update using the UTE μ -map and the T1-weighted (T1w) MR image in order to improve convergence towards a solution. Results show that the modified MLAA algorithm improved the estimated emission image compared to standard OSEM/UTE and OSEM/DIXON. In certain regions of the brain, in particular close to the skull and the air cavities, the modified MLAA algorithm generated less error than OSEM/UTE and OSEM/Dixon. The modified MLAA algorithm is able to compute an attenuation µ-map that is slightly more similar to the aligned CT µ-map than the UTE μ-map.

Primary author: BENOIT, Didier (Rigshospitalet)

Co-authors: Mr HANSEN, Adam Espe (Department of Clinical Physiology, Nuclear Medicine and PET; Rigshospitalet; Denmark); Mr REZAEI, Ahmadreza (University of Leuven); Mr LADEFOGED, Claes (Department of Clinical Physiology, Nuclear Medicine and PET, Rigshospitalet Copenhagen, DK); Dr ANDERSEN, Flemming (Rigshospitalet, University of Copenhagen); Dr NUYTS, Johan (University of Leuven); Dr HØJGAARD, Liselotte (Rigshospitalet, University of Copenhagen); Mr KELLER, Sune (Rigshospitalet, University of Copenhagen); Dr HOLM, Søren (Rigshospitalet, University of Copenhagen)

Presenter: BENOIT, Didier (Rigshospitalet)

Session Classification: Session 6 - Whole-body and dedicated organ MR-PET

Track Classification: 6 - Whole-body and dedicated organ MR-PET