Type: Talk

## Improved Parameter-Estimation with Combined PET-MRI Kinetic Modelling

Monday, 18 May 2015 16:50 (15 minutes)

Kinetic analysis can be applied both to dynamic PET and dynamic contrast enhanced (DCE) MRI data. We have investigated the potential of combined PET-MRI kinetic modelling using simulated FDG data.

The volume of distribution, Ve, for the extra-vascular extra-cellular space (EES) can be estimated by DCE-MRI, and used to reduce the number of parameters in the PET model. We use a 3 tissue-compartment model with 5 rate constants (3TC/5k), in order to distinguish between EES and the intra-cellular space (EIS). In the standard models, k3 represents transfer from the un-metabolised to the metabolised (M) extra-vascular compartment. In our new model, k3'represents transfer from EES to the EIS M-compartment. We also define the more biologically relevant constant, k3"=Vek3', to be used together with the true EES tracer-concentration.

Time-activity curves were generated using the 3TC/5k-model with 3 different Ve-values, but constant k3". Noise was added and the data were fitted with the 2TC/3k model and with the constrained and-un-constrained 3TC/5k model. 100 noise-realisations were generated at 4 different noise-levels.

For the standard 2TC/3k-model, the estimated k3-values were in the range [0.053, 0.094] with SD in the range [0.002,0.043] /min. For the un-constrained 3TC/5k-model, the k3"-values were in the range [0.041,0.187] and SD in [0.053,0.208] /min. With fixed Ve the range of k3" is reduced to [0.083,0.091] with SD in [0.002,0.017] /min. The true k3" value was 0.091/min.

By incorporating information from DCE-MRI into the PET kinetic model, we obtained a good estimate of the parameter k3", independent of Ve.

**Primary author:** ERLANDSSON, Kjell (Institute of Nuclear Medicine, University College London, London, UK)

**Co-authors:** Prof. HUTTON, Brian (Institute of Nuclear Medicine, University College London, London, UK); Dr ATKINSON, David (Centre for Medical Imaging, University College London, London UK); Dr LILJEROTH, Maria (Institute of Nuclear Medicine, University College London, London, UK); Prof. OURSELIN, Sebastien (Centre for Medical Imaging Computing, University College London, London, UK); Prof. ARRIDGE, Simon (Dept of Computer Science, University College London, UK)

Presenter: ERLANDSSON, Kjell (Institute of Nuclear Medicine, University College London, London, UK)

**Session Classification:** Session 3 - Advances in MR-PET and MR-SPECT Software and Quantification

Track Classification: 3 - Advances in MR-PET and MR-SPECT software and quantification