



Contribution ID: 91

Type: Poster

## Building and Commissioning of a Setup to Study Ageing Phenomena in Gaseous Detectors

*Wednesday, 27 May 2015 09:16 (0 minutes)*

In high-rate heavy-ion experiments, gaseous detectors encounter big challenges in terms of degradation of their performance due to a phenomenon called ageing. A setup for high precision ageing studies has been constructed and commissioned at the GSI detector laboratory. Several improvements of the setup design and the gas system have been implemented to achieve the required accuracy of the ageing measurements. The ageing effects of two materials have been carried out: RTV-3145 and Gerband 705. The details of construction of this automatic ageing setup and the test results will be presented.

**Primary author:** Mr ABUHOZA, alhussain (GSI)

**Co-authors:** Dr SCHIMDT, Christian J. (GSI); Prof. SCHMIDT, Hans Rudolf (Eberhard-Karls-Universität, Tübingen, Germany); Mr HEHNER, Joerg (GSI); Dr BISWAS, Saikat (National Institute of Science Education and Research); Dr FRANKENFELD, Ulrich (GSI)

**Presenters:** Dr BISWAS, Saikat (National Institute of Science Education and Research); Mr ABUHOZA, alhussain (GSI)

**Session Classification:** Gas Detectors - Poster Session

**Track Classification:** S7 - Gas detectors